



Calibration Report



RptNo. : ECR2429893

2 of 2

Temperature Item : PT-100(385)

Measured Value (°C)	Actual Value (°C)	Deviation (°C)
-4.65	-5.0	0.4
0.35	0.0	0.4
5.35	5.0	0.4
10.35	10.0	0.4

Calibration Explanation :

- 1.Deviation = Measured – Actual
 - 2.Measured : The value recorded under testing from item- tested.
 - 3.Actual : The value indicated by traceable standard.
 4. The expanded measurement uncertainties estimated are at a level of confidence of approximately 95 % with a coverage factor $k = 2$
 - 5.Expanded uncertainty : 0.6 °C
 - 6.The expanded uncertainty included measurement system variances and item-tested variance.
- THE END --

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.